


Certificate of Analysis

Kit Name: 17 α 20 β Dihydroxyprogesterone ELISA

Components	Expiration Date
17 α 20 β Dihydroxyprogesterone standard 0 ng/mL (0.5 mL)	2025-05
17 α 20 β Dihydroxyprogesterone standard 0.5 ng/mL (0.5 mL)	2025-05
17 α 20 β Dihydroxyprogesterone standard 1.0 ng/mL (0.5 mL)	2025-05
17 α 20 β Dihydroxyprogesterone standard 2.5 ng/mL (0.5 mL)	2025-05
17 α 20 β Dihydroxyprogesterone standard 5.0 ng/mL (0.5 mL)	2025-05
17 α 20 β Dihydroxyprogesterone standard 10.0 ng/mL (0.5 mL)	2025-05
17 α 20 β Dihydroxyprogesterone standard 20.0 ng/mL (0.5 mL)	2025-05
17 α 20 β Dihydroxyprogesterone standard 40.0 ng/mL (0.5 mL)	2025-05
17 α 20 β Dihydroxyprogesterone-HRP Enzyme Conjugate (6.0 mL)	2025-05
17 α 20 β Dihydroxyprogesterone Antibody (2.5 mL)	2025-05
Second Antibody Coated 96 Wells Plate	2025-05
Substrate Solution (12.0mL)	2025-05
Stop Solution (12.0mL)	2025-05



17 α 20 β Dihydroxyprogesterone standard	OD 450 nm
0.5 ng/mL	2.70
1.0 ng/mL	2.43
2.5 ng/mL	1.87
5.0 ng/mL	1.32
10.0 ng/mL	1.02
20.0 ng/mL	0.75
40.0 ng/mL	0.42

This standard curve was obtained on the date of performance testing of the above components. It is for the purpose of illustration only. Each laboratory should establish its own standard curve.

Pass/Reject

Signature with date

Tulsi Immuno Nano Research and Development (TINRAD), LLP, Mandori, Palwal,
Haryana-121102 Email: info@tinrad.com