


## Certificate of Analysis

Kit Name: Cortisol ELISA

Components	Expiration Date
Cortisol standard 0 µg/dL (0.5 mL)	2025-05
Cortisol standard 0.5 µg/dL (0.5 mL)	2025-05
Cortisol standard 1.0 µg/dL (0.5 mL)	2025-05
Cortisol standard 2.5 µg/dL (0.5 mL)	2025-05
Cortisol standard 5.0 µg/dL (0.5 mL)	2025-05
Cortisol standard 10.0 µg/dL (0.5 mL)	2025-05
Cortisol standard 30.0 µg/dL (0.5 mL)	2025-05
Cortisol standard 60.0 µg/dL (0.5 mL)	2025-05
Cortisol-HRP Enzyme Conjugate (6.0 mL)	2025-05
Cortisol Antibody (2.5 mL)	2025-05
Second Antibody Coated 96 Wells Plate	2025-05
Substrate Solution (12.0mL)	2025-05
Stop Solution (12.0mL)	2025-05



Cortisol standard	OD 450 nm
0.5 µg/dL	2.13
1.0 µg/dL	1.88
2.5 µg/dL	1.44
5.0 µg/dL	1.23
10.0 µg/dL	0.88
30.0 µg/dL	0.63
60.0 µg/dL	0.46

This standard curve was obtained on the date of performance testing of the above components. It is for the purpose of illustration only. Each laboratory should establish its own standard curve.

Pass/Reject

Signature with date

Tulsi Immuno Nano Research and Development (TINRAD), LLP, Mandori, Palwal,  
Haryana-121102 Email: [info@tinrad.com](mailto:info@tinrad.com)