


Certificate of Analysis

Kit Name: Testosterone ELISA

Components	Expiration Date
Testosterone standard 0 ng/mL (0.5 mL)	2025-07
Testosterone standard 0.5 ng/mL (0.5 mL)	2025-07
Testosterone standard 1.0 ng/mL (0.5 mL)	2025-07
Testosterone standard 2.5 ng/mL (0.5 mL)	2025-07
Testosterone standard 5.0 ng/mL (0.5 mL)	2025-07
Testosterone standard 10.0 ng/mL (0.5 mL)	2025-07
Testosterone standard 20.0 ng/mL (0.5 mL)	2025-07
Testosterone standard 40.0 ng/mL (0.5 mL)	2025-07
Testosterone-HRP Enzyme Conjugate (6.0 mL)	2025-07
Testosterone Antibody (2.5 mL)	2025-07
Second Antibody Coated 96 Wells Plate	2025-07
Substrate Solution (12.0mL)	2025-07
Stop Solution (12.0mL)	2025-07



Testosterone standard	OD 450 nm
0.5 ng/mL	2.23
1.0 ng/mL	1.99
2.5 ng/mL	1.54
5.0 ng/mL	1.23
10.0 ng/mL	0.88
20.0 ng/mL	0.63
40.0 ng/mL	0.46

This standard curve was obtained on the date of performance testing of the above components. It is for the purpose of illustration only. Each laboratory should establish its own standard curve.

Pass/Reject

Signature with date

Tulsi Immuno Nano Research and Development (TINRAD), LLP, Mandori, Palwal,
Haryana-121102 Email: info@tinrad.com